



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Washington Calibration, LLC
1725 West 3rd Street
Tempe, AZ 85281

Fulfils the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 02 February 2027

Certificate Number: L2152-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Washington Calibration, LLC

1725 West 3rd Street

Tempe, AZ 85281

Rey Feliz

480-820-0506

CALIBRATION

Valid to: February 2, 2027

Certificate Number: L2152-1

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source/Measure	(0 to 100) mV	0.007 mV	Measurement/Comparison to Martel M2000A Calibrator with HP 3458A 8.5 Digit Multimeter
Thermocouple Temperature Simulation Source and Measure	Type J (-300 to 0) °F (0 to 2 100) °F Type K (-300 to 0) °F (0 to 2 400) °F Type N (-430 to 0) °F (0 to 2 370) °F Type R (32 to 2 400) °F Type S (32 to 2 400) °F Type T (-300 to 0) °F (0 to 750) °F	0.51 °F 0.27 °F 0.73 °F 0.44 °F 0.57 °F 0.47 °F 2.4 °F 2.4 °F 0.79 °F 0.38 °F	Measurement/Comparison to Martel M2000A Calibrator with HP 3458A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermocouple Temperature Simulation Controllers & Recorders ¹	Type J (-300 to 0) °F (0 to 2 100) °F Type K (-300 to 0) °F (0 to 2 400) °F Type N (-430 to 0) °F (0 to 2 370) °F Type R (32 to 2 400) °F Type S (32 to 2 400) °F Type T (-300 to 0) °F (0 to 750) °F	0.90 °F 0.79 °F 1.1 °F 0.93 °F 1.0 °F 0.94 °F 2.8 °F 2.8 °F 1.1 °F 0.83 °F	Measurement/Comparison to PIE Thermocouple Calibrator

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Gage Blocks	(0.01 to 4) in	(4.3 + 1L) µin	Comparison to Gage Blocks, Edmunds Gage Block Comparator
	(5 to 20) in	(10 + 1.3L) µin	
Height Masters	(0.5 to 100) mm	(0.11 + 0.001L) µmm	Comparison to Grade 1 Gage Blocks, Labmaster
	(0.01 to 4) in	(6.6 + 0.83L) µin	
Length Standards	Up to 40 in	(39 + 1.3L) µin	Comparison to Reference Bar Gage Blocks
Length Standards	(0.05 to 10) in	(25 + 1.7L) µin	Measurement using Supermicrometer
	(11 to 60) in	(38 + 1.3L) µin	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Feeler Gages (Leaf-Style)	(0.001 to 0.5) in	28 μ in	Measurement using Supermicrometer
Ring Gages	(0.125 to 11) in	(7.2 + 3.4L) μ in	Comparison to Grade 1 Gage Blocks, Labmaster
Plain Plug & Pin Gages	(0 to 10) in	(20 + 2.4L) μ in	Measurement using Supermicrometer
Pin Gages Class Z & ZZ	0 to 0.9 in	54 μ in	Measurement using Laser Micrometer
Thread Plugs	Major Diameter (0.06 to 6) in Pitch Diameter (0.06 to 6) in	(23 + 11L) μ in (81 + 6.4L) μ in	Measurement using Supermicrometer Gage Blocks
Thread Measuring Wires	Unified 60° (4 to 80) TPI Acme 29° (1 to 20) TPI	28 μ in	Measurement using Supermicrometer Gage Blocks
Caliper Masters	Up to 48 in	(39 + 1.3L) μ in	Measurement using Gage Blocks Reference Bar
Surface Plates ¹ Flatness Repeatability	Up to 14 ft x 14 ft ± 0.001 in	(16 + 0.87X) μ in 31 μ in	Measurement using Autocollimator Repeat-Gage
Surface Roughness Standards	(15 to 130) μ in	4.5 μ in	Measurement using Profilometer
2 Pt. Bore Gages	Up to 5 in	63 μ in	Comparison to Supermicrometer
3 Pt. Bore Gages (0.000 1 Resolution) (0.000 2 Resolution) (0.000 5 Resolution)	Up to 5.5 in	73 μ in 190 μ in 290 μ in	Comparison to Ring Gages
Calipers ¹ (0.001 Resolution) (0.0005 Resolution)	(0 to 120) in (0 to 60) in	(580 + 0.2L) μ in (290 + 0.3L) μ in	Comparison to Gage Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Indicators ¹ (0.001 Resolution) (0.000 5 Resolution) (0.000 1 Resolution) (0.000 05 Resolution)	(0 to 4) in	580 μ in 290 μ in 63 μ in 39 μ in	Measurement using Supermicrometer
	(0 to 2) in	13 μ in	
	(0 to 0.1) in	11 μ in	
Supermicrometer ¹ Linearity	(0 to 2) in	11 μ in	Comparison to Gage Blocks
Micrometers, Outside ¹ (0.001 Resolution) (0.000 1 Resolution) (0.000 05 Resolution)	(0 to 40) in	580 μ in $(58 + 1L) \mu$ in	Comparison to Gage Blocks Force Gage Optical Flats Gage Blocks Optical Flats
	(0.5 to 120) in	$(29 + 1.9L) \mu$ in	
Micrometers, Inside ¹ (0.001 Resolution)	(0.5 to 120) in	$(580 + 0.6L) \mu$ in	Comparison to Gage Blocks
Micrometers, Depth ¹ (0.001 Resolution) (0.000 1 Resolution) (0.000 05 Resolution)	(0 to 12) in	580 μ in $(71 + 0.6L) \mu$ in $(46 + 1.1L) \mu$ in	Comparison to Gage Blocks Reference Bar Gage Blocks
	(0 to 40) in	$(580 + 0.2L) \mu$ in $(290 + 0.3L) \mu$ in	
Profilometers	(0.1 μ in Resolution) (1 μ in Resolution)	3.1 μ in 3.1 μ in	Comparison to SRM Reference Patches
Steel Rules	(0 to 72) in	$(280 + 3L) \mu$ in	Measurement using Optical Comparator
	(0 to 72) in	2 900 μ in	Comparison to Standard Rule & Optical Loupe
Tape Measures	(0 to 40) ft	$(2 900 + 0.28L) \mu$ in	Comparison to Standard Rule & Optical Loupe

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Optical Comparators ¹ Magnification	10X, 20X 31.25X, 50X, 62.5X, 100X	170 μ in	Comparison to Precision Ball Standard Magnification Overlay
	(0 to 12) in	91 μ in	Comparison to Stage Micrometer
	0° to 360°	0.01°	Comparison to Steel Square

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment																																								
Rockwell Hardness Testers ¹	HRA HRBW HRC HREW HRFW HRH HR15N	<table> <tbody> <tr> <td>< 70</td> <td>0.33 HRA</td> </tr> <tr> <td>≥ 70 and < 80</td> <td>0.49 HRA</td> </tr> <tr> <td>≥ 80</td> <td>0.19 HRA</td> </tr> <tr> <td>< 60</td> <td>0.40 HRBW</td> </tr> <tr> <td>≥ 60 and < 88</td> <td>0.51 HRBW</td> </tr> <tr> <td>≥ 88</td> <td>0.45 HRBW</td> </tr> <tr> <td>< 35</td> <td>0.41 HRC</td> </tr> <tr> <td>≥ 35 and < 60</td> <td>0.36 HRC</td> </tr> <tr> <td>≥ 60</td> <td>0.37 HRC</td> </tr> <tr> <td>< 84</td> <td>0.51 HREW</td> </tr> <tr> <td>≥ 84 and < 93</td> <td>0.26 HREW</td> </tr> <tr> <td>≥ 93</td> <td>0.53 HREW</td> </tr> <tr> <td>< 80</td> <td>0.52 HRFW</td> </tr> <tr> <td>≥ 80 and < 93</td> <td>0.48 HRFW</td> </tr> <tr> <td>≥ 93</td> <td>0.50 HRFW</td> </tr> <tr> <td>< 96</td> <td>0.44 HRH</td> </tr> <tr> <td>≥ 96</td> <td>0.53 HRH</td> </tr> <tr> <td>< 78</td> <td>0.52 HR15N</td> </tr> <tr> <td>≥ 78 and < 90</td> <td>0.46 HR15N</td> </tr> <tr> <td>≥ 90</td> <td>0.55 HR15N</td> </tr> </tbody> </table>	< 70	0.33 HRA	≥ 70 and < 80	0.49 HRA	≥ 80	0.19 HRA	< 60	0.40 HRBW	≥ 60 and < 88	0.51 HRBW	≥ 88	0.45 HRBW	< 35	0.41 HRC	≥ 35 and < 60	0.36 HRC	≥ 60	0.37 HRC	< 84	0.51 HREW	≥ 84 and < 93	0.26 HREW	≥ 93	0.53 HREW	< 80	0.52 HRFW	≥ 80 and < 93	0.48 HRFW	≥ 93	0.50 HRFW	< 96	0.44 HRH	≥ 96	0.53 HRH	< 78	0.52 HR15N	≥ 78 and < 90	0.46 HR15N	≥ 90	0.55 HR15N	Indirect Verification Method per ASTM E18
< 70	0.33 HRA																																										
≥ 70 and < 80	0.49 HRA																																										
≥ 80	0.19 HRA																																										
< 60	0.40 HRBW																																										
≥ 60 and < 88	0.51 HRBW																																										
≥ 88	0.45 HRBW																																										
< 35	0.41 HRC																																										
≥ 35 and < 60	0.36 HRC																																										
≥ 60	0.37 HRC																																										
< 84	0.51 HREW																																										
≥ 84 and < 93	0.26 HREW																																										
≥ 93	0.53 HREW																																										
< 80	0.52 HRFW																																										
≥ 80 and < 93	0.48 HRFW																																										
≥ 93	0.50 HRFW																																										
< 96	0.44 HRH																																										
≥ 96	0.53 HRH																																										
< 78	0.52 HR15N																																										
≥ 78 and < 90	0.46 HR15N																																										
≥ 90	0.55 HR15N																																										

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers ¹	HR30N < 55 ≥ 55 and < 77 ≥ 77 HR45N < 37 ≥ 37 and < 66 ≥ 66 HR15TW < 81 ≥ 81 and < 87 ≥ 87 HR30TW < 57 ≥ 57 and < 70 ≥ 70 HR15YW < 93 ≥ 93	0.32 HR30N 0.31 HR30N 0.28 HR30N 0.71 HR45N 0.63 HR45N 0.26 HR45N 1.0 HR15TW 0.49 HR15TW 0.50 HR15TW 0.60 HR30TW 0.58 HR30TW 0.21 HR30TW 0.55 HR15YW 0.50 HR15YW	Indirect Verification Method per ASTM E18
Microhardness Testers Indirect Verification of Microhardness Testers ¹	Knoop	HK 300 grf HK 200 grf HK 100 grf	24 HK 19 HK 19 HK
Vickers		HV 500 grf HV 200 grf HV 100 grf	16 HV 19 HV 32 HV

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Class F Weights	25 kg	290 mg	Double Substitution Method – Tolerances per NIST Handbook 105-1
	10 kg	130 mg	
	5 kg	58 mg	
	3 kg	35 mg	
	2 kg	24 mg	
	1 kg	13 mg	
	500 g	8.2 mg	
	300 g	6.7 mg	
	200 g	2.3 mg	
	100 g	1.2 mg	
	50 g	0.69 mg	
	30 g	0.52 mg	
	20 g	0.4 mg	
	10 g	0.29 mg	
	5 g	0.21 mg	
	3 g	0.17 mg	
	2 g	0.15 mg	
	1 g	0.12 mg	
	500 mg	0.093 mg	
	300 mg	0.081 mg	
	200 mg	0.07 mg	
	100 mg	0.058 mg	
	50 mg	0.049 mg	
	30 mg	0.044 mg	
	20 mg	0.041 mg	
	10 mg	0.035 mg	
	5 mg	0.033 mg	
	3 mg	0.029 mg	
	2 mg	0.029 mg	
	1 mg	0.029 mg	

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Class F Weights	50 lb	570 μ lb	Double Substitution Method – Tolerances per NIST Handbook 105-1
	30 lb	330 μ lb	
	25 lb	310 μ lb	
	20 lb	330 μ lb	
	10 lb	140 μ lb	
	5 lb	62 μ lb	
	3 lb	35 μ lb	
	2 lb	31 μ lb	
	1 lb	19 μ lb	
	0.5 lb / 8 oz	6.6 μ lb	
	0.3 lb	4.8 μ lb	
	0.2 lb	3.5 μ lb	
	0.1 lb	2.2 μ lb	
	0.05 lb	1.4 μ lb	
	0.03 lb	1.3 μ lb	
	0.02 lb	0.99 μ lb	
	0.01 lb	0.77 μ lb	
	0.005 lb	0.072 μ lb	
	0.003 lb	0.053 μ lb	
	0.002 lb	0.037 μ lb	
	0.001 lb	0.021 μ lb	
	4 oz	44 μ oz	
	2 oz	29 μ oz	
	1 oz	28 μ oz	
	0.5 oz	22 μ oz	
	0.3 oz	19 μ oz	
	0.25 oz	18 μ oz	
	0.2 oz	16 μ oz	
	0.125 oz	13 μ oz	
	0.1 oz	9.5 μ oz	
	0.062 5 oz	11 μ oz	
	0.05 oz	9.2 μ oz	
	0.031 25 oz	6 μ oz	
	0.03 oz	8.1 μ oz	
	0.015 oz	4.9 μ oz	
	0.02 oz	3 μ oz	
	0.01 oz	3 μ oz	

Mass & Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Torque Wrenches	(4 to 50) ozf·in (25 to 1 000) lbf·in (25 to 1 000) lbf·ft	0.2 ozf·in $0.0076 + 0.0033 \text{ lbf}\cdot\text{in/lbf}\cdot\text{ft}$ $0.36 + 0.0029 \text{ lbf}\cdot\text{ft/lbf}\cdot\text{ft}$	Comparison to Torque Calibration System
Pneumatic Vacuum – Measuring Equipment	(-26 to 100) inHg	$0.005 + 0.037\%$ of reading	Comparison to Mensor 2101Digital Pressure Gauge
Pneumatic Pressure – Measuring Equipment	(0 to 500) psig	0.059 psig	Comparison to Mensor Corp CPC6050 Modular Pressure Controller
Hydraulic Pressure – Measuring Equipment	(0 to 1500) psig	$0.0073 + 0.026\%$ of reading	Comparison to Fluke RPM4-E-DWT A100Me-L Electronic Deadweight Tester


Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Temperature Uniformity Survey ¹	(32 to 2 400) °F (0 to 1 316) °C	($2.5 + 0.0045F$) °F ($1.4 + 0.0045C$) °C	Measurement using Multi-Channel Datalogger with Type K Thermocouple as Per AMS 2750G

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = length in inches or mm where applicable, X = diagonal length in inches, F = temperature reading in Fahrenheit, C = temperature reading in Celsius.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. L2152-1.



Jason Stine, Vice President